

OFFICE OF THE CHIEF DISTRICT MEDICAL & PUBLIC HEALTH OFFICER, KALAHANDI.

Letter No 3014

Dated 24.5.22

To,

The DIO NIC,
Kalahandi.

CORRIGENDUM

As per Tender Notification No. CDM&PHO/KLD/2997/2022-23 on dated 20.05.2022 and subsequent corrigendum issued vide Letter No. 3011 on dated 24.05.2022 towards Tender for Equipment & Instrument for CDM&PHO, Kalahandi for the Year 2022-23 following corrigendum is hereby made to the Tender Technical Specification and earlier corrigendum.

Sl No.	Name of the Item	Tender Technical Specification	Amendment to the Technical Specification
5	BLOOD CELL COUNTER (3PARTS)	Technical Specification: PERFORMANCE PARAMETERS: 2. MID% Analysis: No 4. Gran% Analysis: No 5. Gran# Analysis: No	Technical Specification: Measurement parameters should be RBC, PLT, WBC, HGB, HCT, MCV, MCH, MCHC, RDW-SD /RDW- CV, MPV, LYM%, LYM#, MID%, MID#, GRAN%, GRAN#.

The above corrigendum may be uploaded in the District Website i.e. www.kalahandi.nic.in.

Yours Faithfully

[Signature]
24.5.22

Chief District Medical & Public Health Officer
Kalahandi

24/5/22