

TENDER FOR SUPPLY OF INSTRUMENT, EQUIPMENT &
CONSUMABLES FOR COVID CARE HOME FOR
THE CDM&PHO KALAHANDI

Name of the District / Health Institution: CDM&PHO, KALAHANDI
(HEALTH & F.W. DEPT. ,GOVT. OF ODISHA)

Bid Reference No. CDM&PHO/KLD/ 3375/2020-21


LAST DATE & TIME OF RECEIPT OF BID DOCUMENTS : 22-7-20 at 04.00 PM

DATE & TIME OF OPENING OF THE TENDER : 23-7-20 at 05.00 PM

PLACE OF OPENING OF BID DOCUMENTS
AND
ADDRESS FOR COMMUNICATION
&
RECEIPT OF BID DOCUMENTS

O/o Chief District Medical &PHO,
Kalahandi, Odisha, PIN-766001

Terms and Conditions – Page No. 02 to 03
Instrument & Equipment Name & Specification – Page No. 04 to 06


Chief District Medical &PHO
Kalahandi

**TERMS, CONDITIONS & SPECIFICATION FOR SUPPLY OF
INSTRUMENT, EQUIPMENT & CONSUMABLES FOR COVID CARE
HOME FOR THE CDM&PHO KALAHANDI**

1. The bidders will download the tender from the District Web Site: www.kalahandi.nic.in by Tender cost of Rs.2000/- (Rupees Two thousand only) and EMD Cost. Rs.5000/-(Rupees Five Thousand only) in shape of Demand Draft in favour of ADMO (Medical) Central Store, Kalahandi.
2. The tender paper will be rejected if the bidder changes any clause or Annexure of the bid documents downloaded from the website.
3. Sealed tenders will be received by date 22-7-20 up to 4.00 pm by the office of the CDM&PHO, Kalahandi for the purchase of Instrument, Equipment & Consumable for Covid Care Home. Any tender paper received after the due date and time will be rejected/returned to the sender unopened. The tender paper will be received through Speed Post/ Regd. Post Only.
4. The CDM&PHO shall have no responsibility for any delay / submission on part of the bidder and reserves the right to reject any or all the tenders without assigning any reason thereof.
5. The bidder(s) are to submit their tenders in separate sealed covered envelopes for Technical Bid and Price Bid by superscribing Cover 'A' (Technical Bid) and Cover 'B' (Price Bid) & both the sealed covers should be put into a third outer cover which should be superscribed as "Tender for Supply of Instrument, Equipment & Consumables for Covid Care Home 2020-21" and Tender Reference No. CDM&PHO/ KLD/ 3375 / 2020-21 on dated 14-7-20.
6. The sealed tenders will be opened by the purchase committee in the office chamber of the CDM&PHO, Kalahandi on Dated 22-7-20 at 05.00 pm. The tenderer or their duly authorized representatives are allowed to be present during the opening of the tenders if they so like.
7. The rate quoted by the bidder should be inclusive of all taxes (VAT/CST/ET/Frieght/ Insurance etc)
8. Conditional tenders are liable to be rejected .All Disputes are subjected to Hon'ble Court of Bhawanipatna, Kalahandi, Odisha.
9. If the successful bidder/ bidders fail to supply within the stipulated period i.e. 30 (Thirty) days from date of receipt of final proof from CDM&PHO, Kalahandi liquidated damage @ 0.5% of the tender value per week of delay shall be deducted from the final payment. Maximum delay time acceptable is 8 weeks. Hence, the maximum liquidated damage shall be up to 4% of purchase order. If the bidder fails to supply within the maximum delay time his order stands cancelled automatically.
10. The authority will not make any advance payment to the organization. The organization will have to carry out the entire job on its own and the amount will be paid only after satisfactory completion of the job and submission of bill in that regards.
11. The supply of items shall be made immediately according to volume after placing the supply order in the office of the CDM&PHO, Kalahandi and supplier shall submit the bill for payment at the approved rate in respect to quantity of items supplied. The transportation of items is sole responsibility of the supplier and must deliver the item on door delivery basis.
12. In case of failure on the part of the approved supplier to supply of the above mentioned items as per supply order with stipulated period, the CDM&PHO, Kalahandi shall be at liberty to purchase above mentioned items from other sources and the approved supplier shall be

liable to pay the excess amount which this office have to incur being the different of actual amount of purchase minus the amount as per approved rates and difference aforesaid shall be recoverable and adjustable against the security deposit amount.


13. Under no circumstances shall the organization appoint any sub contractor or sublease the contract. If it is found that the organization has violated these conditions the contract will be terminated forthwith without any notice and security deposited by the organization shall be forfeited.

14. Rates quoted against this tender notice shall remain valid up to 12 months after award of contract. No request of increase in rates, if any, will be allowed or entertained during this period.

15. Bidders should submit the photocopy of valid GST Registration Certificate & PAN Card.

16. Bidders should one year audit report audited by any Chartered accountant and his turn over must be 50 lakh or above

17. Receiving & Opening of Tender may be changed if required by the under Signed, it will be intimated.


Chief District Medical Officer
Kalahandi

INSTRUMENT, EQUIPMENT & CONSUMABLES NAME & SPECIFICATION

1. INFRARED THERMOMETER:-

Specification :-

Product Description : Hand Held Digital Infrared Medical thermometer for measuring human body temperature.

Type of Thermometer : Non contact Infrared Thermometer

Temperature display unit : Degree Celsius & Degree Fahrenheit

Measuring site : Forehead

Measuring Range : $33^{\circ}\text{C} \pm 1^{\circ}\text{C}$ (91.4°F) or Lower to $43^{\circ}\text{C} \pm 1^{\circ}\text{C}$ (109.4°F)

Accuracy of Measurement : 0.2 degree Celsius/ Fahrenheit or better

Display Resolution : 0.1°C/ 0.1°F

Measuring Distance : 5cm or better

Display Type : LCD or LED Display

Features : Auto Shut down when not in use, Audible/ Visual alarm facility

Product Certification : Product should be USFDA approved with 510K number and CE Certified issued from Notified Body.

Manufacturer Standard : ISO 13485 Certified

Warranty : Free Replacement Warranty for 1 year

2. FINGER TYPE PULSE OXIMETER :-

Specification :-

Its should be compact design allows it to be used for adult applications in many different areas such as hospital, homecare and EMS.

It should be with a new, high-tech, colour display known as an 'organic' LED (OLED). It should provide real-time data on your heart rate and blood-oxygen saturation level (SpO2).

In 6 seconds, the oximeter shuts down when you remove your finger with the auto-power off feature. The

- Small, compact, simple, reliable and durable physiological monitoring device.
- Low power consumption with battery-low indicator.
- Dual color OLED displays SpO2, PR, pulse bar and waveform.
- Optional multi-color silicon rubber boot and carrying case.
- Automatically power off and requires 2 AAA alkaline batteries.
- It should be battery powered.

3. BP Instrument (Digital) :-

Specification :-

Type of SPHYGMOMANOMETERS or Blood Pressure recording unit : Digital

Conformity to Indian Standard for SPHYGMOMANOMETERS : NA for Digital

Measuring Device : Automated

Type : Integrated

ISI Marked : NA for Digital

Measurement Method : Oscillometric

Display for Digital BP Instrument : LCD (Liquid Crystal Digital) Display

Range : Pressure Measurement (mmHg) : 0 to 300

Range : Pulse Measurement for Digital with accuracy & Plusmn; 5% (per minute) 40 to 200

Accuracy : Pressure Measurement (mmHg) : +/- 3

Pressure Direction : Capacitive Pressure Sensor

Inflation Automatic pressure Application by Pump Deflation and rapid Air Release : Automatic Pressure Release Valve

Cuff Range : Medium (22-32)cm

Cuff : Disinfect able one piece and bladder cuffs, tested according to ISO 105 EO1/equivalent standard.

Operating Temperature Range : 0 to 55 degree Celsius

Storage Temperature Range : (-10) to 55 degree Celsius

Automatic Power off if System is idle for 03 Minutes : Yes

Weight (in Gram) ; 230

Power

Power Source : AA size chargeable battery set

Type of AA Size Chargeable Battery (AA Size) : Alkaline/LI-ION

4. Glucometer With Strips :-

Specification :-

Product Eligibility Criteria:

- The device should be CE as per IVD or USFDA Certified and the auto disable lancet should be CE or USFDA certified.
- Manufacturer should be ISO 13485 certified for quality standards.
- The test strips should be certified by the National Institutes of Biological (NIB), Govt. of India.

Technical Specifications

1. Small, portable and user friendly device is required. Blood should not go into the glucometer while measurement.
2. It should be able to measure whole blood in capillary mode.
3. Measurement range: 30 to 600 in mg/dl.
4. Accuracy should be as per International Standard ISO 15197 (Requirements for Blood-glucose monitoring systems for self-testing in managing diabetes mellitus). Supporting certificate or test reports from the National Institutes of Biological (NIB) must be furnished of last 2years with the technical bid.
5. Reproducibility/Precision: $\pm 5\%$
6. Display should be 40mm ± 5 mm or better measured diagonally.
7. It should be battery operated electronic system and the battery life should be for atleast 500 tests.
8. Shelf life of strips: Minimum 6 months at the time of delivery to consignee.
9. The test strip should be auto coded or code free.
10. Packing of strips should not be more than 50 strips in a pack.
11. Strips should work for minimum 3 months after opening of strips pack.
12. Operating temperature for both device and test strip should be 10°C to 40°C.
13. Should have strip ejection button facility to avoid cross contamination to the operator.
14. Control solution for checking reliability of strips will be supplied free of cost as & when required.
15. Ready availability of reagent test strips, battery & other consumables across Odisha for at least 5 years.

5. Catgut Chromic Atraumatic Size. 1

6. Catgut Chromic Atraumatic Size. 1.0

7. Catgut Chromic Atraumatic Size. 2.0

8. Catgut Chromic Atraumatic Size. 3.0

9. Polyglactin Size 1

10. Polyglactin Size 1.0


Chief District Medical &PHO
Kalahandi