

**OFFICE OF THE REGULATED MARKET COMMITTEE,  
BHAWANIPATNA**

No. 320

Dated. 27-11-2020

**QUOTATION CALL NOTICE**

Regulated Market Committee, Bhawanipatna invite Sealed quotation as per term condition given from the Dealer/ Authorized Dealers/ OEMS for supply of 30 nos. of Weighing Scales of 300 Kg Capacity and 20 nos. of digital moisture meter for smooth operation of procurement during KMS 2020-21. The specification in pursuant to the guidelines issued by the OSAM Board vide letter No. 1729 dated. 20.08.2019 related to weighing scale and moisture meter is enclosed herewith. The quotation must reach the Office of the Regulated Market Committee, Bhawanipatna by Regd Post / Speed Post/ Courier on or before 7/12/2020 at 1.00 PM and the quotation shall be opened on 08/12/2020 at 11.00 AM at Sub-Collector Office Bhawanipatna in presence of the quotationer or their representative. The quotation shall be finalized by the purchasing committee formed for this purpose. A sum of Rs. 5000.00 only must be deposited as earnest money in the form of Demand Draft in favour of Regulated Market Committee, Bhawanipatna failing which the quotation shall not be considered.


The quotationer must submit papers like VAT, PAN and other related documents as per the norms fixed by the state Govt. and Central Govt. for the purpose of quotation. The equipments should be from a reputed brand and the Farms/ Suppliers having experience with supply of these equipments to RMCs in the state for last 3 year. The payment shall be released after supply of the equipment and in case of Weighing Scale after certification by the Legal Metrology Department. The equipments should be supplied within 10 days from issue of supply order. The Sub-Collector, Bhawanipatna-cum-Chairman, RMC, Bhawanipatna will have the right to reject all or any of the quotation without assigning any reason and to negotiate with any or all the quotationer. The RMC office may be contacted in this regards during office hours.

**Grading and Electronic equipments**


Sl No.	Name of the Equipments	Specification	Quantity
1.	Weighing Scales	As per specification approved by the OSAM Board and copy enclosed herewith.	30
2	Digital Moisture Meter	As per specification approved by the OSAM Board and copy enclosed herewith.	15

Memo No. 321

Office Notice Board, RMC, Bhawanipatna.

  
CHAIRMAN  
RMC, Bhawanipatna


Dated. 27-11-2020

  
CHAIRMAN  
RMC, Bhawanipatna

Memo No. 322

Dated. 27-11-2020

Copy along with the term & condition forwarded to the District Information Officer, Kalahandi for information and he is requested to host the quotation call notice in the Kalahandi district portal from 27-11-2020 to 07-12-2020 upto 3 P.M. for its wider publicity.


  
CHAIRMAN  
RMC, Bhawanipatna

**TECHNICAL SPECIFICATION OF ELECTRONIC WEIGHING SCALE 300Kg.  
CAPACITY (With Printer facility)**

Sl. No.	Specification
1	Minimum weight of the machine 30Kg.
2	Weighing Capacity – 300 Kg. maximum, 1 Kg minimum e=50 g.(least count)
3	Accuracy classes – Class-III
4	Platform size – 60 cm x 60 cm
5	Platform-Powder coated frame and stainless steel platter (202 grade & above)
6	Indicator / Display – LED/LCD display screen
7	In case of printer facility, the machine should have computer interface and printing facility with memory in the printer.
8	Load Cell : The load cell must be overload and shock protection. Connecting wires from base to head should be rat protection.
9	Battery – Minimum 12 hours & above Battery Backup. Battery housing should be separated and should be replaced without breaking seal of Inspector Legal Metrology.
10	With both software and hardware locking facility. Provision for seal of Inspector Legal Metrology
11	Overload 600 Kg. & above
12	Model approval : Weighing Scale have model approval certificate from competent authority from Govt. of India
13	Machines shall be delivered after certification by LMO.

**TECHNICAL SPECIFICATION OF DIGITAL MOISTURE METER**

Sl. No.	Specifications
1	Measuring Range: 3.5% to 40%
2	Supply Voltage : 9V DC
3	Temperature Compensation : Automatic calibration & error correction facility
4	Display : Digital LCD / LED
5	Error condition display for over Range or under range : $\pm 0.2\%$
6	Commodity : Cereals, Pulses, Oil Seeds
7	Operating Temperature : Ambient to 50° C

  
**Chairman**  
**R.M.C. Bhawanipal**